

Class 120 KISTOCK KT 120 and KH 120



Table of contents

1	Safety instructions.....	4
1.1	Precautions for use.....	4
1.2	Symbols used.....	4
2	Presentation of the device.....	5
2.1	Use.....	5
2.2	Applications.....	5
2.3	Description of the device.....	5
2.4	Description of the keys.....	5
2.5	PC connection.....	5
2.6	Fixation.....	6
3	Technical features.....	6
3.1	Devices.....	6
3.2	Housing.....	7
3.3	Dimensions.....	7
3.4	Guarantee period.....	7
4	Use of the device.....	8
4.1	Display.....	8
4.2	Functions of keys.....	8
4.2.1	Groups organisation.....	10
4.3	Datalogger configuration with the integrated PDF file.....	10
4.4	Datalogger download with PDF report edition.....	13
4.5	Configuration, datalogger download and data processing with the KILOG software.....	14
5	Device maintenance.....	15
5.1	Replace the battery.....	15
5.2	Device cleaning.....	15
6	Calibration.....	16
7	Accessories.....	16
8	Troubleshooting.....	16

1 Safety instructions

1.1 Precautions for use

Please always use the device in accordance with its intended use and within parameters described in the technical features in order not to compromise the protection ensured by the device.

1.2 Symbols used

For your safety and in order to avoid any damage of the device, please follow the procedure described in this user manual and read carefully the notes preceded by the following symbol:



The following symbol will also be used in this user manual:
Please read carefully the information notes indicated after this symbol.



2 Presentation of the device

2.1 Use

The KT 120 and KH 120 dataloggers of the HVAC range allow the internal measurement of temperature only (KT 120) or of temperature and humidity (KH 120). This class of devices is dedicated to the food transport.

The devices have a male USB plug and an integrated software in PDF format which enables to download and configure the datalogger without specific software.

2.2 Applications

The KISTOCK datalogger is ideal for a temperature and humidity control for the sensitive foodstuff storage, for example in the food industry or pharmaceutical domain. It allows to control the temperature and humidity in refrigerators, cold rooms, food trucks etc.

Therefore, the device guarantees a traceability all along the cold chain. And at any moment the KISTOCK datalogger allows to edit easily and quickly a data report in PDF format.




2.3 Description of the device



2.4 Description of the keys

 "OK" key: allows to validate, start or stop the records, display the value


 "Selection" key: allows the functions scroll

2.5 PC connection



2.6 Fixation

The KT 120 and KH 120 KISTOCK dataloggers have a magnetic mounting, so you can fix it easily.



3 Technical features

3.1 Devices

	KT 120	KH 120
Units displayed	°C, °F	°C, °F, %RH
Resolution	0.1 °C, 0.1 °F	0.1 °C, 0.1 °F, 0.1 %RH
External input	USB connector	
Internal sensor	Temperature	Temperature, humidity
Type of sensor	NTC	Temperature: NTC Humidity: capacitive
Measuring range	From -40 to +70 °C	Temperature: From -20 to +70 °C Humidity: From 0 to 100 %RH
Accuracies*	±0.4 °C (-20 °C < T < +70 °C) ±0.8 °C (beyond)	Temperature: From -20 to 0 °C: ±2 % of the reading value ± 0.6 °C From 0 to 30 °C: ± 0.5 °C From 30 to 70 °C: ± 1.5 % of the reading value Humidity: Accuracy (Repeatability, linearity, hysteresis): ±2 %RH (from 15 °C to 25 °C, from 5 to 95 %RH) Factory calibration uncertainty: ±0.88 %RH Temperature dependence: ±0.04 x (T-20) %RH (if T ≤ 15 °C or T ≥ 25 °C)
Setpoint alarms	2 setpoint alarms on each channel	
Number of points	50 000	
Frequency of measurement	From 1 minute to 24 hours	
Working temperature	From -40 to +70 °C	From -20 to +70 °C
Storage temperature	From -40 to +85 °C	
Battery life	3 years**	500 days**
European directives	2011/65/EU RoHS II ; 2012/19/EU WEEE ; 2004/108/EC EMC ; 2006/95/EC	


* All accuracies indicated in this document were stated in laboratory conditions and can be guaranteed for measurement carried out in the same conditions, or carried out with calibration compensation.

** On the basis of 1 measurement each 15 minutes at 25 °C

3.2 Housing

Dimensions	100 x 42.5 x 15.9 mm
Weight	53 g
Display	1-line LCD screen Dimensions of screen: 32 x 25.5 mm
Control	1 OK key 1 Selection key
Material	Compatible with food industry environment ABS housing
Protection	IP65: KT 120 IP40: KH 120
PC communication	1 USB A male input
Battery power supply	1 x CR2450 (button battery)
Environmental conditions of use	Air and neutral gases Humidity: in non condensing conditions Maximum altitude: 2000 m


3.3 Dimensions



3.4 Guarantee period

KISTOCK dataloggers have 1-year guarantee for any manufacturing defect (return to our After-sales service required).

4.1 Display



END DATASET is finished

REC Indicates that one value is being recorded. It flashes: the DATASET did not start already.

FULL Flashing slowly: DATASET is between 80 and 90 % of the storage capacity.
Flashing quickly: DATASET is between 90 and 100 % of the storage capacity.
Constant: storage capacity full.

BAT Constant: indicates that the batteries have to be replaced.

12 Indicates the channel number which is measuring.

°C Temperature in °C.

MIN The displayed values are the recorded

°F Temperature in °F.

MAX maximum/minimum values for the displayed channels.

Indicates the alarm action type: rising or falling action.

%RH Relative humidity (KH 120).

The values to display selected during configuration via the software will scroll on the screen every 3 seconds (only with the KH 120).

The display can be activated or deactivated via the KILOG software.

At extreme temperatures, the display can become hardly readable and its display speed can slow down at temperatures below 0 °C. This has no incidence on the measurement accuracy.

4.2 Functions of keys

OK key: enables to start, stop the dataset (press during >3 seconds) or to change of scrolling group as described in the tables below.

Selection key: enables the scroll values in the scrolling group as described in the tables below.

Device state	Type of start/stop selected	Key used	Action generated	Illustration
Waiting for start flashes	Start: by key	 During 3 seconds	Dataset starting	
	Stop: indifferent		Inactive	
	Start by PC or date/hour		Inactive	
	Stop: indifferent		Inactive	
	Start: indifferent		Measurements scrolling (group 1)*	<p>Or wait 3 seconds</p>
	Stop: indifferent			

* Please see the summary table of the groups organization on page 10.

** Only with the KH 120.


Device state	Type of start/stop selected	Key used	Action generated	Illustration
Dataset in progress REC	Start: indifferent Stop: by key	 During 3 seconds	Dataset stop	
	Start: indifferent Stop: indifferent		Group change (groups 2 and 3)*	
	Start: indifferent Stop: indifferent		Groups scrolling (groups 1, 2 and 3)*	
Dataset finished END	Indifferent		Inactive	
	Indifferent		Measurements scrolling*	<p>Or wait 3 seconds</p>

* Please see the summary table of the groups organization on the following page.

** Only with the KH 120.


4.2.1 Groups organisation

The table below summarises the groups organisation and measured values available during a measurement dataset.



The diagram shows a table with three columns: Group 1, Group 2, and Group 3. Above the table, two 'OK' keys are shown with arrows pointing to the column headers. To the left of the table, a 'Refresh' key is shown with an arrow pointing to the first row of the table.

	Group 1	Group 2	Group 3
Measured temperature		Max. value in temperature Min. value in temperature	High alarm threshold in temperature Low alarm threshold in temperature
Measured hygrometry*		Max. value in hygrometry Min. value in hygrometry	High alarm threshold in hygrometry Low alarm threshold in hygrometry

Press  key to change of group.

Press  key to scroll the values in the group.

4.3 Datalogger configuration with the integrated PDF file


The class 120 KISTOCK dataloggers have an integrated PDF file which allows to configure quickly and easily the datalogger. Therefore, you can directly configure your datalogger without opening the KILOG software.



Required configuration: to open this document, you need to use ONLY the “**Adobe Acrobat Reader 9**®” program (or higher), freely downloadable, which allows to read PDF format documents. Ensure you have installed it before starting.

- Plug the class 120 KISTOCK datalogger on an USB port of your computer**.
The following window opens:



- Click on “**Open folder to view files**”.
Wait a few seconds (according to the dataset number of points), and a volume appears.
- Double-click on the “**Configuration...**” PDF file  configuration KT [1K 15.05.99999]

* Only with the KH 120.

** The computer must be in compliance with the IEC60950 standard.

The integrated configuration file opens:


The file **header** indicates the device name and serial number.

Button which allows to **load a previous configuration**, saved before on the computer.

“General information”
Displays the name of the dataset and the possible comments.

“Configuration” allows to configure:
- The datalogger
- The recording mode
- The stop conditions of the record

“Channel parameters” allows to configure the channels of temperature (KT 120 and KH 120) and hygrometry (only with the KH 120).




- **Choose the language**

Select the required language:



- **Choose the date format**

Select the required date format:



- **General information**

Dataset name: this field allows to name the dataset.

Comments: this field allows to write comments on dataset.

General informations



Dataset:

Comments:

- **Configuration**

Recorder

Active screen: tick “Yes” to activate the screen display or “No” to deactivate it.

Management DST: for an automatic management of DST, tick “Yes” or tick “No” to deactivate it. If you choose “Yes”, the “Next time change” fields become accessible. Dates and times of the next time change are proposed by default. You can modify them: click on the “Date” field then on  to display the calendar. Click on the required date. Click on the “Hour” field to modify the time on which the next time change will be applied: the time format is 00:00. On the last field, click on  and choose “+1h” to add an hour or “-1h” to subtract an hour. The date and time change will be applied on the required date and time and will add or subtract an hour.



Configuration

Recorder

Active screen Yes No

Management DST Yes No

Next time change

Date: 01/03/2016 02:00 +1h


Date: 30/10/2016 03:00 -1h

Recording

Interval: in the “Interval” field, inform the required interval duration between two measurements, then select the unit (minutes or hours).

Start type

- Tick “Button” for a start type by button.

- Tick “Date” for a start type by date: inform the required date and time of the start. Click on the “Start date” field then click on  to display the calendar and select the required date, or write it manually, with respect to the date format selected previously.

Recording

Interval: 10 Minutes

Start Type: Button Date

Start date: 20/10/2015 12:00


Stop conditions

Stop type

Select the required stop type:

Stop type

- Duration
- Date
- Duration
- Nbr of items
- Total Memory
- Loop

- The stop by date is available only if the start type by date has been chosen previously. If you choose “Date”, inform the required stop date and time in the “Stop date” field: click on  to display the calendar then select the required date, or write it manually with respect to the date format selected previously.

- The stop by duration allows to determine a recording duration: inform the “Days” and “Hours” fields.

- The stop by number of items allows to determine a required number of measurement before the dataset stops. Inform the “Number of items” field (between 1 and 50 000 points).

- “Total memory” allows to record continuously up to 50 000 points before the dataset stops.

- “Loop” allows to record the values continuously and once the memory capacity reached, the last recorded values overwrite the first.

- “Stop by button”: tick “Yes” to allow a stop by button. Therefore, push the datalogger OK key during 3 seconds to stop the measurement dataset. To not allow it, tick “No”. The stop by button can not be deactivated if the chosen stop condition is “Loop”, “Total memory” or “Number of points”.



Stop conditions

Stop type: Duration

Stop date: 18/11/2015 12:00

No. of items: 4 321

Duration: Days 30 Hours 00:00

Stop by button: Yes No

Channel parameters

Temperature (KT 120 and KH 120) and humidity (only the KH 120)

- Choose the temperature measurement unit: tick the “°C” or “°F” box.

- Active channel: tick “Yes” to activate the channel or “No” to deactivate it.

- Active alarm: tick “Yes” to activate the alarm or “No” to deactivate it.

Active channel Yes No

Active alarm Yes No

Channel parameters

Temperature °C °F

Active channel Yes No

Active alarm Yes No

If the alarm is activated, inform the “High” and “Low” fields to configure the alarm threshold. Inform the “Delay” field in number of points. According to the measurement interval previously configured, the duration is automatically actualised. For example: if an interval of 1 minute has been configured and that the delay in number of points for the high threshold is 5, the delay duration will be 5 minutes.

High Low

Alarm threshold: 80,0 -20,0

Delay (nb pts): 5 0



Delay (Duration): 00:05 00:00

- When the configuration is finished, click on the  button to validate.
- Chose the place to save the configuration: to use this configuration for the next dataset, save the configuration directly on the datalogger on “**Removable disk**”.



Check that the dataset is finished to ensure the new configuration is taken into account.

A message asks you to overwrite the existing file.


- Click on “**Yes**”.
- To use this configuration later for another dataset, or to configure another device, you can save it on the required location. To recover it later, click on the  button on top of the page, then select the required file on “.xdp” format, click on the  button to use it for the next dataset.

4.4 Datalogger download with PDF report edition

- Plug the class 120 KISTOCK datalogger on an USB plug of the computer*.







Wait a few seconds, then the following window opens:

- Click on “**Open folder to view files**”.
- The windows explorer opens.*
- Double-click on the “**Report**” PDF file to visualise the dataset report. 

* The computer must be in compliance with the IEC60950 standard.

Example of report:


State indicators:

-  : Alarm activated
-  : Alarm deactivated
-  : Alarm OK
-  : Alarm triggered

The **background information** indicates the dataset name and the eventual comments.

The **device configuration** is summarised with the measurement interval, the start/stop type, the recording duration, the number of recorded values and the date and time of dataset start and end.

The **chart** allows to visualise the recorded values evolution, with the units on ordinate, the date and time on abscissa and the defined high and low thresholds.



The **header** indicates the date and time of the report edition

Device references Serial number and firmware version

The **channel statistics** indicate the values unit and the minimal, maximal, average and MKT values.

The **alarms parameters and results** indicate the defined high and low thresholds, their state (alarm or OK) and the longest duration of threshold exceeding.

➤ You can print it or export it in PDF format to integrate it easily to your documents.

4.5 Configuration, datalogger download and data processing with the KILOG software

Please see the KILOG software user manual: “KILOG-classes-50-120-220-320”.


5 Device maintenance

5.1 Replace the battery

With 500 days to 3 years* battery life, KISTOCK guarantees long-term measurement.

To replace the battery:

1. Unlock the battery hatch with a screwdriver or a coin.
2. Turn towards the left until the marker aligns in front of the opened padlock symbol.
3. Continue to turn until the hatch rises.
4. Replace the battery (button battery CR 2450**) in such a way the + pole will be visible.



- Replace the battery hatch with the indicator in front of the opened padlock and close it by turning it towards the right in order to make correspond the indicator with the closed padlock.



Only use trademark or high quality batteries in order to guarantee the announced autonomy.



After the battery replacement, the device must be reconfigured.

5.2 Device cleaning

Please avoid any aggressive solvent.

Please protect the device from any cleaning produce containing formalin, that may be used for cleaning rooms and ducts.

* On the basis of 1 measurement each 15 minutes at 25 °C

** The battery must be in compliance with the 60086-4 standard.




6 Calibration

All the KISTOCK devices have an integrated adjustment certificate in the memory in PDF format which can be visualized and printed easily.

A calibration certificate is available as option in paper format.

We recommend to carry out a yearly checking.

7 Accessories

Accessories	Part numbers	Illustrations
1 button battery CR2450	KBL-2450	
KILOG software	KILOG-3-N	
Calibration certificate	-	-
25 mm diameter metal washer with double sided adhesive tape	KRM	



Only the accessories supplied with the device must be used.

8 Troubleshooting

Problem	Probable cause and possible solution
"hi" or "lo" is displayed.	The measurement range is exceeded. There is a problem with the sensing element.
No value is displayed, only the icons are present.	The display is configured on "OFF". Configure it on "ON" with the KILOG software (see page 14).
The display is completely off and there is no communication with the computer.	The battery has to be replaced. (see page 15).



BE CAREFUL! Material damages can happen, so please apply the precautionary measures indicated.



Once returned to KIMO, required waste collection will be assured in the respect of the environment in accordance to guidelines relating to WEEE.

www.kimo.fr

EXPORT DEPARTMENT
Boulevard de Beaubourg - Emerainville - BP 48
77312 MARNE LA VALLEE CEDEX 2
Tel : + 33.1.60.06.69.25 - Fax : + 33.1.60.06.69.29

